

NANO M.ON

CERTIFIKÁT FILTRAČNÍ ÚČINNOSTI

výrobce / distributor: **Nano Medical s.r.o.
FERMATA, a.s.**

výrobek: **Prémiová nanovlákná
zdravotnická maska CE**

typ: **NANO M.ON**

NANOVLÁKENNÁ MASKA NANO M.ON JE CERTIFIKOVANÁ
NA **FILTRAČNÍ ÚČINNOST VĚTŠÍ NEŽ 94%** DLE
PŘÍSLUŠNÝCH NOREM.

**Potvrzujeme, že je obdobným prostředkem ochrany dýchacích cest
dle Mimořádného opatření Vlády ČR a Ministerstva zdravotnictví.**

 A Sotera Health company	Sponsor: Marešová Muzarová Nano Medical s.r.o. S května 1640/85 Praha 14000 CZECH REPUBLIC
Latex Particle Challenge Final Report	
Test Article: AntiMicrobe Web R	
Study Number: 122279-001	
Study Received Date: 28 Feb 2020	
Testing Facility: Nelson Laboratories, LLC	
6280 S. Redwood Rd.	
Salt Lake City, UT 84123 U.S.A.	
Test Procedure(s): Standard Test Protocol (STP) Number: STP0005 Rev 07	
Deviation(s): None	
Test Side: Either Side	Area Tested: 91.5 cm ²
Particle Size: 0.1 µm	Laboratory Conditions: 21 °C, 24% relative humidity (RH) at 0654, 21 °C, 23% RH at 1200
Average Filtration Efficiency: 99.7%	Standard Deviation: 0.042
 Study Director: Curtis Gerow, B.S.	 Study Completion Date: 16 Apr 2020
	
865.958.7400 nelsonlabs.com info@nelsonlabs.com	



Jiří Fantík DiS.

člen představenstva FERMATA, a.s.

v Čelakovicích dne 28. 2. 2021

